

برگشت (/Panel/Labs/Answers/47998)

چاپ

دوره پنجاه و سوم زمستان 1404

S2 Results

نمونه مجهول ToRCH

کد نمونه 1173	EQAP-18421	آزمایشگاه آزمایشگاه آرامش
---------------	------------	---------------------------

Anti-CMV (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

Anti-Rubella (IgG)

Anti-Toxoplasma (IgG)

Anti-Toxoplasma (IgG)

جدول شماره 1. میزان مشارکت در این دوره-آزمایش

جدول شماره 2. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت

← مشخص شده است.

تعداد ثبت نام کنندگان	تعداد شرکت کنندگان	کل درصد مشارکت	تعداد پاسخهای دریافتی	درصد پاسخدهی
191	168	% 87.96	160	% 95.24

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	44	42	0	2	%95.45	%0.00	%4.55
ECL	28	26	1	1	%92.86	%3.57	%3.57
ELFA	2	2	0	0	%100.00	%0.00	%0.00
ELISA	78	77	0	1	%98.72	%0.00	%1.28
Immunochromatography	1	1	0	0	%100.00	%0.00	%0.00
Other	6	5	1	0	%83.33	%16.67	%0.00

جدول شماره 3. توزیع فراوانی نتایج کیفی گزارش شده به تفصیل بر اساس کیت و روش

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
	Abbott	Positive	10	
	Architect	Positive	1	
	Dia Sorin	Negative	1	
		Positive	16	
	Liaison	Positive	2	
	Maglumi	Not Reported	1	
		Negative	1	
		Positive	3	
	Mindray	Positive	1	
	Ortho	Positive	1	
	Others	Positive	1	
Roche	Positive	2		
Siemens	Positive	4		
ECL	Not Reported	Not Reported	1	
	Abbott	Positive	1	
	Cobas	Positive	2	
	Liaison	Positive	1	
	Maglumi	Indeterminate	1	
		Positive	1	
	Roche	Positive	21	
Siemens	Negative	1		
ELFA	Chorus	Positive	2	

Method	Kit	Result	Count	Your Answer
ELISA	Not Reported	Positive	2	
	Chorus	Negative	1	
		Positive	3	
	Diesse-Tec	Positive	2	
	Euroimmun	Positive	3	
	Pishtaz Teb	Positive	61	←
	Spinreact	Positive	1	
	Vircell	Positive	5	
Immunochromatography	Maglumi	Positive	1	
Other	Chorus	Indeterminate	1	
		Positive	2	
	Diesse-Tec	Positive	2	
	Maglumi	Positive	1	
RIA	Pishtaz Teb	Positive	1	

Copyright © 2026 IACLD. All rights reserved.