

[برگشت \(/Panel/Labs/Answers/54255\)](#)

چاپ

دوره پنجاه و چهارم تابستان 1405

S2 Results

نمونه مجهول ToRCH

کد نمونه 1146	EQAP-18421	آزمایشگاه آزمایشگاه آرامش
---------------	------------	---------------------------

Anti-CMV (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

Anti-Rubella (IgG)

Anti-Toxoplasma (IgG)

Anti-Toxoplasma (IgG)

جدول شماره 1. میزان مشارکت در این دوره-آزمایش

جدول شماره 2. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت

← مشخص شده است.

تعداد ثبت نام کنندگان	تعداد شرکت کنندگان	کل درصد مشارکت	تعداد پاسخهای دریافتی	درصد پاسخدهی
207	189	% 91.30	173	% 91.53

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	38	35	2	1	%92.11	%5.26	%2.63
ECL	38	35	2	1	%92.11	%5.26	%2.63
ELFA	4	3	1	0	%75.00	%25.00	%0.00
ELISA	83	80	2	1	%96.39	%2.41	%1.20
Immunochromatography	1	1	0	0	%100.00	%0.00	%0.00
Other	9	8	1	0	%88.89	%11.11	%0.00

جدول شماره 3. توزیع فراوانی نتایج کیفی گزارش شده به تفصیل بر اساس کیت و روش

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Not Reported	1	
		Positive	1	
	Abbott	Positive	9	
	Architect	Positive	4	
	Dia Sorin	Not Reported	1	
		Indeterminate	1	
		Positive	7	
	Diaplus	Positive	1	
	Liaison	Negative	1	
		Positive	4	
	Maglumi	Indeterminate	1	
		Positive	4	
	Ortho	Positive	1	
	Roche	Positive	1	
	Siemens	Positive	3	
ECL	Abbott	Positive	2	
	Chorus	Indeterminate	1	
	Cobas	Positive	4	
	Dia Sorin	Negative	1	
		Positive	2	
	Liaison	Positive	2	
	Roche	Positive	25	
	Siemens	Indeterminate	1	
ELFA	Not Reported	Positive	1	
	Abbott	Positive	1	←
		Indeterminate	1	
		Positive	1	

Chorus

Method	Kit	Result	Count	Your Answer
ELISA		Indeterminate	2	
		Negative	1	
		Positive	6	
	Euroimmun	Positive	1	
	Pishtaz Teb	Positive	69	
	Vircell	Positive	4	
Immunochromatography	Siemens	Positive	1	
Other	Architect	Positive	1	
	Chorus	Positive	4	
	Dia Sorin	Positive	1	
	Diesse-Tec	Indeterminate	1	
		Positive	1	
	Pishtaz Teb	Not Reported	1	
		Positive	1	