

برگشت (/Panel/Labs/Answers/47996)

چاپ

دوره پنجاه و یکم تابستان 1404

S1.1 Results

نمونه مجهول ایمونوسرولوژی برای آزمایش HBsAg, HIVAb, HCVAb

کد نمونه 2488	EQAP-18421	آزمایشگاه آزمایشگاه آرامش
---------------	------------	---------------------------

S1/1 HBs Ag

S1/1 HCV Ab

S1/1 HIV Ab

S1/1 HCV Ab

جدول شماره 1. میزان مشارکت در این دوره-آزمایش

جدول شماره 2. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ — بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت

← مشخص شده است.

تعداد ثبت نام کنندگان	تعداد شرکت کنندگان	کل درصد مشارکت	تعداد پاسخهای دریافتی	درصد پاسخدهی
1683	1604	% 95.31	1499	% 93.45

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	55	0	0	55	%0.00	%0.00	%100.00
ECL	139	1	0	138	%0.72	%0.00	%99.28
ELFA	12	0	0	12	%0.00	%0.00	%100.00
ELISA	1267	6	1	1260	%0.47	%0.08	%99.45
Immunochromatography	18	1	0	17	%5.56	%0.00	%94.44
Other	7	0	0	7	%0.00	%0.00	%100.00

جدول شماره 3. توزیع فراوانی نتایج کیفی گزارش شده به تفصیل بر اساس کیت و روش

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Negative	3	
	Abbott	Negative	27	
	Architect	Negative	3	
	Mindray	Negative	9	
	Ortho	Negative	4	
	Roche	Negative	1	
	Siemens	Negative	8	
ECL	Not Reported	Negative	1	
	Abbott	Negative	8	
	Architect	Negative	2	
	Cobas	Negative	10	
		Positive	1	
	Mindray	Negative	3	
	Roche	Not Reported	1	
		Negative	110	
	Siemens	Negative	4	
ELFA	Biomerieux	Negative	11	
	Biorex	Negative	1	

Method	Kit	Result	Count	Your Answer
ELISA	Not Reported	Not Reported	1	
		Indeterminate	1	
		Negative	48	
	AB Germany	Negative	2	
	Abbott	Negative	1	
	Ai De	Negative	1	
	Arya Mabna Tashkhis	Negative	51	
		Positive	2	
	Bahar Afshan	Negative	1	
	Biomerieux	Negative	2	
	Bio-Pharm	Negative	1	
	Diapro	Negative	19	
	Euroimmun	Negative	1	
	Gbc	Not Reported	1	
		Negative	104	
		Positive	1	
	General Biologicals(GB)	Negative	29	←
	Mindray	Negative	2	
	Monobind	Negative	2	
	Others	Negative	2	
	Padtan Teb	Negative	7	
	Pgi	Negative	1	
	Pishtaz Teb	Not Reported	1	
		Negative	927	
		Positive	3	
	Sam Fan Avaran	Negative	1	
	Spinreact	Negative	2	
	Usa By W.H.P.M	Negative	1	
	Wantai	Negative	55	

Method	Kit	Result	Count	Your Answer
Immunochromatography	Not Reported	Negative	6	
	Abbott	Negative	1	
	Abon	Negative	5	
	Acon	Positive	1	
	ImmunoDiagnostic Systems	Negative	1	
	Liaison	Negative	1	
	Pishtaz Teb	Negative	1	
	Siemens	Negative	1	
	Vircell	Negative	1	
Other	Not Reported	Negative	2	
	Abon	Negative	1	
	Architect	Negative	1	
	Biomerieux	Negative	2	
	Ortho	Negative	1	

Copyright © 2025 IACLD. All rights reserved.