

برگشت (/Panel/Labs/Answers/47996)

چاپ

دوره پنجاه و یکم تابستان 1404

S2 Results

نمونه مجهول ToRCH

کد نمونه 1175	EQAP-18421	آزمایشگاه آزمایشگاه آرامش
---------------	------------	---------------------------

Anti-CMV (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

Anti-Rubella (IgG)

Anti-Toxoplasma (IgG)

Anti-CMV (IgG)

جدول شماره 1. میزان مشارکت در این دوره-آزمایش

جدول شماره 2. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت

← مشخص شده است.

تعداد ثبت نام کنندگان	تعداد شرکت کنندگان	کل درصد مشارکت	تعداد پاسخهای دریافتی	درصد پاسخدهی
196	187	% 95.41	167	% 89.30

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	53	52	0	1	%98.11	%0.00	%1.89
ECL	32	32	0	0	%100.00	%0.00	%0.00
ELFA	3	3	0	0	%100.00	%0.00	%0.00
ELISA	74	71	0	3	%95.95	%0.00	%4.05
Immunochromatography	1	1	0	0	%100.00	%0.00	%0.00
Other	4	4	0	0	%100.00	%0.00	%0.00

جدول شماره 3. توزیع فراوانی نتایج کیفی گزارش شده به تفصیل بر اساس کیت و روش

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
	Abbott	Positive	8	←
	Architect	Positive	4	
	Dia Sorin	Positive	15	
	Liaison	Positive	7	
	Maglumi	Positive	6	
	Ortho	Positive	1	
	Roche	Positive	2	
	Siemens	Negative	1	
		Positive	8	
ECL	Abbott	Positive	1	
	Architect	Positive	1	
	Cobas	Positive	3	
	Dia Sorin	Positive	1	
	Liaison	Positive	3	
	Roche	Positive	20	
	Siemens	Positive	3	
ELFA	Biomerieux	Positive	2	
	Chorus	Positive	1	
ELISA	Not Reported	Positive	2	
	Chorus	Negative	1	
		Positive	4	
	Diesse-Tec	Negative	1	
	Euroimmun	Positive	2	
	Mindray	Positive	1	
	Padtan Teb	Positive	1	
	Pishtaz Teb	Positive	59	
	Vircell	Negative	1	
		Positive	2	

Method	Kit	Result	Count	Your Answer
Immunochromatography	Chorus	Positive	1	
Other		Positive	2	
	Diesse-Tec	Positive	2	

Copyright © 2025 IACLD. All rights reserved.