

[برگشت \(/Panel/Labs/Answers/47996\)](#)

چاپ

دوره پنجاه و یکم تابستان 1404

S2 Results

نمونه مجهول ToRCH

کد نمونه 1175	EQAP-18421	آزمایشگاه آزمایشگاه آرامش
---------------	------------	---------------------------

Anti-CMV (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

Anti-Rubella (IgG)

Anti-Toxoplasma (IgG)

Anti-Rubella (IgG)

جدول شماره 1. میزان مشارکت در این دوره-آزمایش

جدول شماره 2. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت

← مشخص شده است.

تعداد ثبت نام کنندگان	تعداد شرکت کنندگان	کل درصد مشارکت	تعداد پاسخهای دریافتی	درصد پاسخدهی
196	187	% 95.41	181	% 96.79

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	51	51	0	0	%100.00	%0.00	%0.00
ECL	28	28	0	0	%100.00	%0.00	%0.00
ELFA	5	5	0	0	%100.00	%0.00	%0.00
ELISA	90	89	0	1	%98.89	%0.00	%1.11
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	5	5	0	0	%100.00	%0.00	%0.00

جدول شماره 3. توزیع فراوانی نتایج کیفی گزارش شده به تفصیل بر اساس کیت و روش

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
	Abbott	Positive	13	←
	Architect	Positive	4	
	Dia Sorin	Positive	15	
	Liaison	Positive	6	
	Maglumi	Positive	5	
	Ortho	Positive	1	
	Roche	Positive	1	
	Siemens	Positive	5	
ECL	Not Reported	Not Reported	1	
	Abbott	Positive	2	
	Architect	Positive	2	
	Cobas	Positive	1	
	Dia Sorin	Positive	1	
	Liaison	Positive	3	
	Maglumi	Positive	1	
	Roche	Positive	17	
	Siemens	Positive	1	
ELFA	Biomerieux	Positive	2	
	Chorus	Positive	2	
	Dia Sorin	Positive	1	
ELISA	Not Reported	Positive	4	
	Chorus	Positive	3	
	Others	Positive	1	
	Pishtaz Teb	Not Reported	1	
		Negative	1	
		Positive	73	
	Vircell	Positive	8	

Method	Kit	Result	Count	Your Answer
Other	Abbott	Positive	1	
	Chorus	Positive	1	
	Diesse-Tec	Positive	3	
RIA	Pishtaz Teb	Positive	1	

Copyright © 2025 IACLD. All rights reserved.